

Bulletin CL420-EN

www.yokogawa.com/ymi/







# CL420 Clamp-on Process Meter

#### Product Model Code


Name	Model
Clamp-on Process Meter	CL420

#### Standard Accessories (supplied)

Name	Model
Soft case	93045

#### Optional Accessories (sold separately)

Name	Model
Output cable (banana plug)	98076
Output cable (for screw terminal)	98077



# **Specifications**

Model	CL420		
Diameter of measurable conductor	6mm diameter max.		
DC Current	Range and resolution [Range	Accuracy <sup>⁻1</sup>	
	20 mA: 0.00 to±21.49mA	±(0.2%rdg+5dgt) <sup>2</sup>	
	100 mA: ±21.0 to ±126.0mA	±(1.0%rdg+5dgt)	
DC Voltage OUTPUT [10mV/mA]	20 mA: 0.0 to±214.9mV	(DCA Accuracy)+(±0.5mV)	
	100 mA: ±210 to±1260mV	(DCA Accuracy)+(±3mV)	
Display	4-digit LCD Numeric display		
Response time	Approx. 1.5 seconds (2.5 seconds when across the range)		
Range switching	Auto range		
Operating temperature and humidity	-10°C to +50°C 80% RH or less (no condensation)		
Safety Standards	EN61010-1, EN61010-2-030, EN61010-2-032		
Withstanding voltage	2.21kV AC for 5 seconds (between the core and the case)		
Power supply	Four AA-size alkaline batteries (1.5V LR6)		
Battery life	Approx. 60hrs (continuous) backlight off and LED light off		
Other functions	Data hold, Zero adjust function, Auto power off, LED Torch		
	light, Back light display, Illuminant panel		
External dimensions and weight	61 (W) × 111 (H) × 40 (D) mm		
	Approx. 290g (including batteries)		
Standard accessories	User's Manual, Batteries, Soft case (93045)		

# **Examples of Analog Output Application**



Example of 98077 output cable for screw terminal connecting to a data logger



#### Accessories (sold separately)



Output cable Model: 98076

Output cable (for screw terminal) Model: 98077



Accessories (supplied)

Soft case for CL420

# **Related Products**



Pressure Calibrator CA700



Multi-functional Calibrator CA150



Multi-functional Calibrator **CA71** 



**Process Multimeter** CA450

#### CAUTION



Read the product manual before using this product to ensure

# YOKOGAWA TEST & MEASUREMENT CORPORATION

Global Sales Dept. /Phone: +81-422-52-6237 Facsimile: +81-422-52-6462 E-mail: tm@cs.jp.yokogawa.com

YOKOGAWA CORPORATION OF AMERICA YOKOGAWA EUROPE B.V. YOKOGAWA TEST & MEASUREMENT (SHANGHAI) CO., LTD. YOKOGAWA ELECTRIC KOREA CO., LTD.

YOKOGAWA ENGINEERING ASIA PTE. LTD. YOKOGAWA INDIA LTD. YOKOGAWA ELECTRIC CIS LTD. YOKOGAWA AMERICA DO SUL LTDA.

YOKOGAWA MIDDLE EAST & AFRICA B.S.C(c)

Phone: +1-800-888-6400

Phone: +31-88-4641429 Phone: +86-21-6239-6363

Facsimile: +86-21-6880-4987 Facsimile: +82-2-2628-3899 Phone: +82-2-2628-3810 Phone: +65-6241-9933 Facsimile: +65-6241-9919 Phone: +91-80-4158-6396 Facsimile: +91-80-2852-1442 Phone: +7-495-737-78-68 Facsimile: +7-495-737-78-69

Phone: +55-11-3513-1300 Phone: +973-17-358100

Facsimile: +973-17-336100

# Represented by:

tmi.yokogawa.com/

YMI-VP-LF-S-EN06



<sup>\*1</sup> At 23°C ±5°C, 75% RH or less
Measurement accuracy: ±(% of reading + digits)
Terms of accuracy: Open and close the clamp sensor after power on and perform zero adjustment. \*2 The 20mA range accuracy assurance is the average of 5 times measuring